

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3351

Name: RITTER2  
Sterilizer ID: V2313228  
02/02/2024 02:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 107.6 F 0.0  
2:00 109.5 F 0.1  
4:00 134.1 F 1.9  
6:00 164.3 F 5.4  
8:00 186.5 F 9.7  
10:00 219.2 F 7.0  
12:00 224.1 F 8.0  
14:00 237.0 F 12.7  
16:00 234.9 F 9.6  
18:00 237.7 F 9.3  
20:00 247.9 F 14.4  
22:00 259.9 F 21.2  
24:00 271.0 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 29.3  
1:00 272.8 F 29.6  
2:00 272.7 F 29.9  
3:00 272.0 F 29.6  
4:00 272.4 F 29.3

Min 271.8 F 29.2  
Max 273.2 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.7 F 0.0  
5:00 222.6 F 0.0  
10:00 199.2 F 0.0  
15:00 186.5 F 0.0  
20:00 182.5 F 0.0  
25:00 182.0 F 0.0  
30:00 180.3 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:12  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:03:28  
CYCLE COMPLETE

02/02/2024 03:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_