

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2242

Name: RITTER2  
Sterilizer ID: V2313228  
04/12/2023 03:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.2 F 0.0  
2:00 115.2 F 0.1  
4:00 143.9 F 2.4  
6:00 173.0 F 6.2  
8:00 193.1 F 10.9  
10:00 210.2 F 4.5  
12:00 224.7 F 9.2  
14:00 233.7 F 10.9  
16:00 238.7 F 11.3  
18:00 237.3 F 9.6  
20:00 250.9 F 16.2  
22:00 262.5 F 23.0  
24:00 272.7 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.8 F 29.6  
2:00 271.6 F 29.2  
3:00 272.5 F 29.6  
4:00 271.5 F 28.9

Min 271.4 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 221.7 F 0.0  
10:00 196.7 F 0.0  
15:00 200.7 F 0.0  
20:00 198.7 F 0.0  
25:00 197.1 F 0.0  
30:00 208.4 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 24:18  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:02:30  
CYCLE COMPLETE

04/12/2023 04:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_