
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000060
17:07 19/01/2021
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CYCLE ABORTED 2:02
STOP BUTTON PRESSED
NOT STERILE
END TIME 2:02

Digital Signature #
FE266032E544134C
