
Midmark M11 - v1.0.5
Total Cycles: 1347

Name: RITTER2
Sterilizer ID: V2313228
07/26/2022 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.6 F 0.0
2:00 194.7 F 1.3
4:00 201.9 F 6.8
6:00 220.5 F 12.0
8:00 227.8 F 8.2
10:00 240.7 F 12.5
12:00 241.2 F 12.1
14:00 241.0 F 11.3
16:00 254.3 F 18.0
18:00 265.6 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.7 F 29.6
2:00 271.7 F 29.0
3:00 272.8 F 29.7
4:00 271.9 F 29.0

Min 271.6 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 205.1 F 0.0
10:00 213.6 F 0.0
15:00 208.6 F 0.0
20:00 204.7 F 0.0
25:00 203.0 F 0.0
30:00 189.4 F 0.0
Drying Complete

FILLING: 2:57
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

07/26/2022 12:15 PM
ACCEPTED: X
REJECTED:
Operator: U02