
Midmark M11 - v1.0.5
Total Cycles: 4300

Name: RITTER2
Sterilizer ID: V2313228
01/27/2025 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.1 F 0.0
2:00 115.5 F 1.1
4:00 133.8 F 3.7
6:00 163.9 F 6.8
8:00 182.1 F 10.2
10:00 198.6 F 9.4
12:00 216.5 F 6.2
14:00 224.9 F 10.0
16:00 231.2 F 10.1
18:00 235.5 F 11.0
20:00 234.4 F 8.8
22:00 246.5 F 14.7
24:00 256.9 F 20.4
26:00 266.5 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.9
1:00 271.4 F 30.5
2:00 273.3 F 31.2
3:00 272.1 F 30.5
4:00 272.4 F 31.3

Min 270.9 F 29.7
Max 273.4 F 31.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 200.1 F 0.0
10:00 194.2 F 0.0
15:00 192.1 F 0.0
20:00 178.4 F 0.0
25:00 177.1 F 0.0
30:00 174.4 F 0.0

Drying Complete

FILLING: 1:00
HEATING: 27:51
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:06:52
CYCLE COMPLETE

01/27/2025 04:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____