

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1366

Name: RITTER 1  
Sterilizer ID: V2178198  
11/15/2021 07:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

| mm:ss | Degrees | PSI  |
|-------|---------|------|
| 0:00  | 70.1 F  | 0.0  |
| 2:00  | 74.2 F  | 0.0  |
| 4:00  | 84.7 F  | 0.0  |
| 6:00  | 114.0 F | 0.1  |
| 8:00  | 141.0 F | 0.2  |
| 10:00 | 170.5 F | 1.1  |
| 12:00 | 188.4 F | 3.4  |
| 14:00 | 199.3 F | 5.4  |
| 16:00 | 214.4 F | 8.0  |
| 18:00 | 228.1 F | 11.9 |
| 20:00 | 227.5 F | 6.1  |
| 22:00 | 236.2 F | 10.5 |
| 24:00 | 235.5 F | 8.6  |
| 26:00 | 244.6 F | 12.7 |
| 28:00 | 241.6 F | 11.1 |
| 30:00 | 253.5 F | 17.1 |
| 32:00 | 263.7 F | 23.1 |
| 34:00 | 272.5 F | 28.7 |

STERILIZING

| mm:ss | Degrees | PSI  |
|-------|---------|------|
| 0:00  | 271.1 F | 28.3 |
| 1:00  | 272.4 F | 29.0 |
| 2:00  | 272.4 F | 28.6 |
| 3:00  | 271.1 F | 28.1 |
| 4:00  | 271.9 F | 28.8 |

Min 271.0 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING

| Drying Start    | mm:ss | Degrees | PSI |
|-----------------|-------|---------|-----|
|                 | 0:00  | 211.2 F | 0.0 |
|                 | 5:00  | 222.1 F | 0.0 |
|                 | 10:00 | 220.8 F | 0.0 |
|                 | 15:00 | 216.2 F | 0.0 |
|                 | 20:00 | 201.7 F | 0.0 |
| Drying Complete | 25:00 | 228.4 F | 0.0 |
|                 | 30:00 | 233.8 F | 0.0 |

FILLING: 1:00  
HEATING: 34:18  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:12:43  
CYCLE COMPLETE

11/15/2021 09:09 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_