

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1405

Name: RITTER2  
Sterilizer ID: V2313228  
08/24/2022 08:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.0 F 0.0  
2:00 179.8 F 1.6  
4:00 197.9 F 4.3  
6:00 218.8 F 9.1  
8:00 221.6 F 9.5  
10:00 229.6 F 9.4  
12:00 237.8 F 9.7  
14:00 241.4 F 12.5  
16:00 240.9 F 11.3  
18:00 253.7 F 17.9  
20:00 264.8 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.1  
1:00 272.6 F 29.8  
2:00 271.6 F 29.0  
3:00 272.5 F 29.9  
4:00 272.1 F 29.3

Min 271.5 F 28.9  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 219.7 F 0.0  
10:00 238.7 F 0.0  
15:00 231.5 F 0.0  
20:00 228.7 F 0.0  
25:00 230.8 F 0.0  
30:00 230.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 21:55  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:00:00  
CYCLE COMPLETE

08/24/2022 09:46 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_