
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007203
9:35 15/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:34
STERILIZING 4:24
136.0 C 220kPa 4:24
136.4 C 225kPa 4:54
136.3 C 222kPa 5:24
136.4 C 222kPa 5:54
136.4 C 223kPa 6:24
136.2 C 224kPa 6:54
136.4 C 225kPa 7:24
136.1 C 217kPa 7:54
136.1 C 217kPa 7:54
VENTING 7:54
AIR DRYING 8:48
CYCLE COMPLETE 9:31

Digital Signature #
20A0B49B35A90F47
