
Midmark M11 - v1.0.0
Total Cycles: 3615

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2023 05:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.3 F 0.0
2:00 163.1 F 0.7
4:00 182.6 F 4.2
6:00 211.8 F 8.9
8:00 221.0 F 9.7
10:00 232.0 F 8.9
12:00 236.5 F 11.3
14:00 240.7 F 11.2
16:00 237.6 F 9.3
18:00 251.3 F 16.0
20:00 262.8 F 22.5
22:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.8
2:00 271.9 F 28.9
3:00 272.1 F 28.6
4:00 272.1 F 29.1

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 231.1 F 0.0
10:00 224.9 F 0.0
15:00 221.1 F 0.0
20:00 228.2 F 0.0
25:00 228.9 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

08/03/2023 06:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____