

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4096

Name: RITTER2  
Sterilizer ID: V2313228  
11/07/2024 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.2 F 0.0  
2:00 172.0 F 1.7  
4:00 175.8 F 5.1  
6:00 194.9 F 8.9  
8:00 210.6 F 13.0  
10:00 216.9 F 6.1  
12:00 229.1 F 10.7  
14:00 232.7 F 7.9  
16:00 242.9 F 12.4  
18:00 244.5 F 13.0  
20:00 256.3 F 19.1  
22:00 266.7 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 272.7 F 30.1  
2:00 272.0 F 29.4  
3:00 272.9 F 30.0  
4:00 271.9 F 29.5

Min 271.0 F 28.9  
Max 273.2 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
5:00 218.7 F 0.0  
10:00 225.6 F 0.0  
15:00 229.8 F 0.0  
20:00 228.7 F 0.0  
25:00 230.0 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 23:38  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:02:56  
CYCLE COMPLETE

11/07/2024 12:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_