
Midmark M11 - v1.0.5
Total Cycles: 252

Name: RITTER2
Sterilizer ID: V2313228
07/27/2021 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.4 F 0.0
2:00 167.4 F 3.0
4:00 187.6 F 7.9
6:00 207.7 F 13.5
8:00 221.9 F 9.7
10:00 234.6 F 9.0
12:00 241.5 F 12.6
14:00 244.1 F 13.0
16:00 255.5 F 19.6
18:00 266.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 30.0
1:00 272.2 F 30.0
2:00 272.7 F 30.3
3:00 271.9 F 30.7
4:00 271.6 F 30.0

Min 271.4 F 29.9
Max 272.8 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 228.6 F 0.0
10:00 229.2 F 0.0
15:00 230.7 F 0.0
20:00 226.9 F 0.0
25:00 217.4 F 0.0
30:00 184.6 F 0.0

Drying Complete

FILLING: 2:02
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

07/27/2021 11:43 AM
ACCEPTED: X
REJECTED:
Operator: U01