
Midmark M11 - v1.0.0
Total Cycles: 1797

Name: RITTER 1
Sterilizer ID: V2178198
04/06/2022 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 74.8 F 0.0
4:00 86.2 F 0.0
6:00 117.7 F 0.1
8:00 154.0 F 1.1
10:00 174.6 F 4.0
12:00 191.9 F 7.5
14:00 212.8 F 11.0
16:00 223.6 F 5.0
18:00 226.4 F 7.8
20:00 238.7 F 11.9
22:00 234.9 F 8.3
24:00 244.9 F 12.2
26:00 244.3 F 12.2
28:00 256.2 F 18.4
30:00 266.6 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.7
2:00 271.2 F 28.2
3:00 272.4 F 28.7
4:00 271.2 F 28.2

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 230.7 F 0.0
10:00 194.4 F 0.0
15:00 185.4 F 0.0
20:00 185.7 F 0.0
25:00 202.0 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:39
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

04/06/2022 09:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____