
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005403
10:28 21/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:55
STERILIZING 2:59
136.4 C 223kPa 2:59
136.4 C 224kPa 3:29
136.1 C 219kPa 3:59
136.2 C 223kPa 4:29
136.3 C 225kPa 4:59
136.4 C 223kPa 5:29
136.2 C 222kPa 5:59
136.2 C 220kPa 6:29
136.2 C 220kPa 6:29
VENTING 6:29
AIR DRYING 7:19
CYCLE COMPLETE 9:24

Digital Signature #
EFADFA16C4C67C6D
