
Midmark M11 - v1.0.0
Total Cycles: 1851

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2022 07:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 74.0 F 0.0
4:00 83.5 F 0.0
6:00 128.0 F 0.3
8:00 162.4 F 2.8
10:00 189.5 F 6.4
12:00 213.9 F 10.3
14:00 217.8 F 7.1
16:00 227.1 F 7.3
18:00 238.8 F 11.6
20:00 232.8 F 7.7
22:00 245.5 F 12.9
24:00 243.4 F 11.7
26:00 256.6 F 18.6
28:00 268.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 272.5 F 28.9
2:00 272.1 F 28.8
3:00 271.7 F 28.5
4:00 272.0 F 28.3

Min 271.0 F 27.9
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 171.9 F 0.0
10:00 165.9 F 0.0
15:00 170.3 F 0.0
20:00 163.5 F 0.0
25:00 161.0 F 0.0
30:00 159.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:05
STERILIZING: 4:00
VENTING: 2:52
DRYING: 30:00
TOTAL CYCLE: 01:06:49
CYCLE COMPLETE

04/21/2022 08:55 AM
ACCEPTED: X
REJECTED:
Operator: U02