
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006981
11:38 26/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:37
STERILIZING 3:26
136.0 C 220kPa 3:26
136.2 C 222kPa 3:56
136.2 C 221kPa 4:26
136.4 C 224kPa 4:56
136.2 C 224kPa 5:26
136.1 C 221kPa 5:56
136.4 C 224kPa 6:26
136.4 C 224kPa 6:56
136.4 C 224kPa 6:56
VENTING 6:56
AIR DRYING 7:45
CYCLE COMPLETE 9:00

Digital Signature #
72D1053433AC9F04
