

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3877

Name: RITTER2  
Sterilizer ID: V2313228  
08/21/2024 09:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.5 F 0.0  
2:00 73.9 F 0.0  
4:00 79.1 F 0.0  
6:00 106.4 F 1.2  
8:00 142.2 F 3.9  
10:00 182.5 F 7.3  
12:00 188.1 F 11.4  
14:00 208.9 F 3.0  
16:00 216.0 F 6.7  
18:00 227.6 F 10.4  
20:00 232.1 F 9.0  
22:00 238.4 F 11.2  
24:00 235.4 F 8.9  
26:00 248.7 F 15.0  
28:00 259.6 F 21.1  
30:00 269.7 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.0  
1:00 272.3 F 29.6  
2:00 272.0 F 29.1  
3:00 272.1 F 29.6  
4:00 272.2 F 29.2

Min 271.2 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
5:00 204.0 F 0.0  
10:00 188.5 F 0.0  
15:00 186.9 F 0.0  
20:00 184.8 F 0.0  
25:00 183.4 F 0.0  
30:00 184.5 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:52  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:09:01  
CYCLE COMPLETE

08/21/2024 10:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_