
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001429
15:09 09/13/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:44
PRESSURIZING 6:29
STERILIZING 8:01
136.0 C 220kPa 8:01
136.2 C 221kPa 8:31
136.4 C 220kPa 9:02
136.4 C 221kPa 9:32
136.4 C 223kPa 10:02
136.2 C 222kPa 10:32
136.4 C 221kPa 11:02
136.4 C 220kPa 11:31
VENTING 11:32
AIR DRYING 12:37
CYCLE COMPLETE 22:43

Digital Signature #
83EBE2383444DDC4
