
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000253
8:09 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:52
STERILIZING 5:28
136.4 C 233kPa 5:28
136.4 C 233kPa 5:58
136.4 C 233kPa 6:28
136.3 C 233kPa 6:58
136.4 C 233kPa 7:28
136.2 C 230kPa 7:58
136.4 C 232kPa 8:28
136.3 C 232kPa 8:58
136.3 C 232kPa 8:58
VENTING 8:58
AIR DRYING 10:01
CYCLE COMPLETE 23:16

Digital Signature #
A374C16F3170C5CE
