

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2948

Name: RITTER 1  
Sterilizer ID: V2178198  
02/27/2023 02:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
2:00 180.7 F 3.9  
4:00 196.5 F 8.3  
6:00 218.0 F 11.2  
8:00 228.6 F 8.7  
10:00 239.3 F 11.0  
12:00 240.6 F 11.7  
14:00 239.3 F 10.2  
16:00 253.1 F 17.1  
18:00 265.2 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.4 F 28.9  
2:00 271.6 F 28.7  
3:00 272.4 F 28.9  
4:00 271.3 F 28.6  
  
Min 271.3 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 217.9 F 0.0  
10:00 233.1 F 0.0  
15:00 232.7 F 0.0  
20:00 225.8 F 0.0  
25:00 224.1 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 19:41  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:26  
CYCLE COMPLETE

02/27/2023 03:25 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_