
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007845
7:35 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:01
PRESSURIZING 4:01
STERILIZING 5:01
136.2 C 220kPa 5:01
136.4 C 221kPa 5:31
136.1 C 219kPa 6:01
136.3 C 223kPa 6:31
136.0 C 218kPa 7:01
136.2 C 220kPa 7:31
136.1 C 221kPa 8:01
136.4 C 218kPa 8:31
136.4 C 218kPa 8:31
VENTING 8:31
AIR DRYING 9:25
CYCLE COMPLETE 10:00

Digital Signature #
8777BF3E46693054
