
Midmark M11 - v1.0.0
Total Cycles: 4992

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2024 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.7 F 0.0
2:00 154.7 F 0.4
4:00 167.8 F 1.2
6:00 196.4 F 5.3
8:00 209.3 F 9.7
10:00 217.8 F 9.0
12:00 227.2 F 8.0
14:00 238.7 F 12.6
16:00 236.8 F 9.6
18:00 238.5 F 9.2
20:00 247.4 F 13.9
22:00 258.8 F 20.1
24:00 269.2 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.7 F 28.9
2:00 271.7 F 28.2
3:00 272.7 F 29.1
4:00 272.3 F 28.6

Min 271.6 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 198.7 F 0.0
10:00 200.8 F 0.0
15:00 201.5 F 0.0
20:00 208.3 F 0.0
25:00 213.8 F 0.0
30:00 214.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:06
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:03:28
CYCLE COMPLETE

10/11/2024 02:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____