
Midmark M11 - v1.0.0
Total Cycles: 4108

Name: RITTER 1
Sterilizer ID: V2178198
12/12/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 71.4 F 0.2
4:00 74.3 F 1.0
6:00 86.9 F 2.8
8:00 125.1 F 5.0
10:00 178.4 F 7.6
12:00 182.1 F 10.8
14:00 216.5 F 6.4
16:00 213.9 F 5.3
18:00 226.8 F 8.6
20:00 236.4 F 12.1
22:00 231.8 F 7.3
24:00 238.5 F 11.9
26:00 236.8 F 9.2
28:00 248.7 F 14.9
30:00 259.0 F 20.6
32:00 268.7 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.1 F 29.1
2:00 271.1 F 28.7
3:00 272.1 F 29.1
4:00 271.1 F 28.5

Min 271.1 F 28.5
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 226.0 F 0.0
10:00 238.9 F 0.0
15:00 227.5 F 0.0
20:00 222.3 F 0.0
25:00 224.9 F 0.0
30:00 229.7 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:11
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:11:45
CYCLE COMPLETE

12/12/2023 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____