
Midmark M11 - v1.0.5
Total Cycles: 2488

Name: RITTER2
Sterilizer ID: V2313228
06/08/2023 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.3 F 0.0
2:00 196.0 F 4.1
4:00 207.7 F 9.9
6:00 222.9 F 6.9
8:00 235.9 F 12.0
10:00 237.1 F 10.3
12:00 236.6 F 9.4
14:00 251.7 F 16.7
16:00 264.0 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.4 F 29.5
2:00 271.7 F 29.3
3:00 272.3 F 29.5
4:00 271.5 F 29.1

Min 271.5 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 216.4 F 0.0
10:00 232.3 F 0.0
15:00 224.9 F 0.0
20:00 221.3 F 0.0
25:00 220.3 F 0.0
30:00 220.3 F 0.0

Drying Complete

FILLING: 2:22
HEATING: 17:56
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:33
CYCLE COMPLETE

06/08/2023 02:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____