
Midmark M11 - v1.0.0
Total Cycles: 2008

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2022 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.6 F 0.0
2:00 160.8 F 3.0
4:00 170.8 F 6.6
6:00 193.2 F 11.2
8:00 214.6 F 5.6
10:00 228.2 F 10.8
12:00 230.4 F 6.8
14:00 240.3 F 12.7
16:00 239.7 F 10.5
18:00 251.7 F 16.8
20:00 263.3 F 23.5
22:00 272.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.7 F 29.7
2:00 272.3 F 29.2
3:00 271.6 F 29.1
4:00 272.9 F 29.8

Min 271.4 F 28.8
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 222.7 F 0.0
10:00 229.1 F 0.0
15:00 230.9 F 0.0
20:00 229.7 F 0.0
25:00 229.0 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

06/08/2022 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____