
Midmark M11 - v1.0.0
Total Cycles: 2196

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2022 03:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.7 F 0.0
2:00 190.1 F 4.2
4:00 199.5 F 9.0
6:00 225.1 F 7.5
8:00 231.8 F 9.1
10:00 239.1 F 10.4
12:00 241.9 F 11.8
14:00 240.2 F 10.3
16:00 253.4 F 16.9
18:00 264.9 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.6
2:00 271.6 F 28.2
3:00 272.7 F 28.7
4:00 271.6 F 28.0

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 201.9 F 0.0
10:00 204.1 F 0.0
15:00 203.9 F 0.0
20:00 202.3 F 0.0
25:00 202.0 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 19:52
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:55
CYCLE COMPLETE

08/03/2022 04:49 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____