
Midmark M11 - v1.0.5
Total Cycles: 1518

Name: RITTER2
Sterilizer ID: V2313228
09/21/2022 05:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.3 F 0.0
2:00 187.3 F 3.9
4:00 200.4 F 8.9
6:00 218.3 F 9.4
8:00 228.5 F 9.8
10:00 233.5 F 8.7
12:00 243.7 F 13.5
14:00 243.5 F 12.6
16:00 256.1 F 19.3
18:00 267.3 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.8 F 30.0
2:00 272.0 F 29.3
3:00 271.3 F 29.3
4:00 272.7 F 30.0

Min 271.2 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 235.3 F 0.0
10:00 227.3 F 0.0
15:00 227.3 F 0.0
20:00 223.4 F 0.0
25:00 222.9 F 0.0
30:00 224.3 F 0.0

Drying Complete

FILLING: 2:16
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

09/21/2022 06:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____