
Midmark M11 - v1.0.5
Total Cycles: 82

Name: RITTER2
Sterilizer ID: V2313228
06/09/2021 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.4 F 0.0
2:00 151.0 F 2.6
4:00 170.9 F 6.7
6:00 199.0 F 12.3
8:00 215.2 F 7.1
10:00 231.0 F 13.0
12:00 233.3 F 10.9
14:00 237.4 F 9.6
16:00 249.3 F 16.6
18:00 261.7 F 23.8
20:00 272.7 F 30.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.2
1:00 272.9 F 30.6
2:00 272.2 F 30.4
3:00 272.8 F 30.6
4:00 272.2 F 30.2

Min 271.8 F 30.0
Max 273.1 F 31.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 231.2 F 0.0
10:00 228.7 F 0.0
15:00 231.5 F 0.0
20:00 225.1 F 0.0
25:00 226.8 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:17
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

06/09/2021 11:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____