
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001474
10:47 10/06/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:17
PRESSURIZING 7:15
STERILIZING 9:15
136.1 C 220kPa 9:15
136.4 C 220kPa 9:45
136.4 C 221kPa 10:15
136.4 C 222kPa 10:45
136.2 C 220kPa 11:15
136.5 C 220kPa 11:45
136.4 C 217kPa 12:15
136.3 C 221kPa 12:45
136.3 C 221kPa 12:45
VENTING 12:45
AIR DRYING 13:49
CYCLE COMPLETE 43:49

Digital Signature #
68D9A2EC6EF8EA40
