
Midmark M11 - v1.0.5
Total Cycles: 113

Name: RITTER2
Sterilizer ID: V2313228
06/17/2021 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.0 F 0.0
2:00 140.1 F 1.6
4:00 144.7 F 4.2
6:00 168.0 F 7.6
8:00 191.3 F 11.6
10:00 204.5 F 4.1
12:00 219.9 F 8.4
14:00 231.6 F 13.0
16:00 234.6 F 10.5
18:00 236.0 F 9.3
20:00 248.3 F 15.2
22:00 259.3 F 21.2
24:00 269.4 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 271.9 F 30.1
2:00 272.8 F 30.2
3:00 272.5 F 29.9
4:00 272.0 F 29.6

Min 271.2 F 29.3
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 212.3 F 0.0
10:00 219.4 F 0.0
15:00 228.1 F 0.0
20:00 222.9 F 0.0
25:00 187.9 F 0.0
30:00 183.8 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 25:17
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:04:13
CYCLE COMPLETE

06/17/2021 01:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____