
Midmark M11 - v1.0.0
Total Cycles: 4755

Name: RITTER 1
Sterilizer ID: V2178198
07/22/2024 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.6 F 0.0
2:00 191.2 F 3.8
4:00 193.1 F 8.2
6:00 211.2 F 13.0
8:00 218.5 F 7.1
10:00 233.2 F 11.8
12:00 233.0 F 8.8
14:00 239.9 F 10.4
16:00 245.9 F 13.3
18:00 257.8 F 19.7
20:00 268.4 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 28.9
2:00 271.5 F 28.8
3:00 272.1 F 28.8
4:00 271.3 F 28.7

Min 271.1 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 221.2 F 0.0
10:00 220.7 F 0.0
15:00 227.8 F 0.0
20:00 230.0 F 0.0
25:00 225.7 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

07/22/2024 01:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____