

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2193

Name: RITTER 1  
Sterilizer ID: V2178198  
08/03/2022 12:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.5 F 0.0  
2:00 124.5 F 0.3  
4:00 150.6 F 3.2  
6:00 176.9 F 6.8  
8:00 199.2 F 11.1  
10:00 211.8 F 4.1  
12:00 224.0 F 8.4  
14:00 235.0 F 12.9  
16:00 233.8 F 9.0  
18:00 240.9 F 11.0  
20:00 243.8 F 12.2  
22:00 255.3 F 18.2  
24:00 265.8 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.2 F 29.2  
2:00 272.3 F 28.9  
3:00 271.5 F 28.3  
4:00 271.5 F 28.7

Min 271.2 F 28.2  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.9 F 0.0  
5:00 200.7 F 0.0  
10:00 223.0 F 0.0  
15:00 225.4 F 0.0  
20:00 226.7 F 0.0  
25:00 229.7 F 0.0  
30:00 233.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 25:54  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:03:56  
CYCLE COMPLETE

08/03/2022 01:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_