
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 007197
12:44 12/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:05
STERILIZING 3:07
136.0 C 220kPa 3:07
136.5 C 225kPa 3:37
136.4 C 225kPa 4:07
136.3 C 222kPa 4:37
136.2 C 222kPa 5:07
136.4 C 224kPa 5:37
136.4 C 224kPa 6:07
136.4 C 224kPa 6:37
136.4 C 224kPa 6:37
VENTING 6:37
AIR DRYING 7:25
CYCLE COMPLETE 37:25

Digital Signature #
79909E77B76D3275
