
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 000844
8:43 20/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:47
STERILIZING 4:54
136.0 C 220kPa 4:54
136.2 C 223kPa 5:24
136.3 C 222kPa 5:54
136.4 C 223kPa 6:24
136.4 C 220kPa 6:54
136.2 C 223kPa 7:24
136.4 C 224kPa 7:54
136.3 C 223kPa 8:24
136.3 C 223kPa 8:24
VENTING 8:24
AIR DRYING 9:18
CYCLE COMPLETE 9:23

Digital Signature #
8DFE35591CDCBDB9
