
Midmark M11 - v1.0.0
Total Cycles: 3871

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2023 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.4 F 0.0
2:00 175.4 F 3.1
4:00 192.7 F 7.0
6:00 208.7 F 12.2
8:00 226.7 F 6.4
10:00 236.1 F 11.5
12:00 234.6 F 8.4
14:00 241.5 F 11.1
16:00 245.8 F 13.0
18:00 257.7 F 19.4
20:00 268.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 271.8 F 28.9
2:00 271.8 F 28.9
3:00 271.8 F 28.7
4:00 271.9 F 28.8

Min 271.5 F 28.7
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 198.6 F 0.0
10:00 199.0 F 0.0
15:00 199.1 F 0.0
20:00 195.8 F 0.0
25:00 192.3 F 0.0
30:00 189.3 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

10/05/2023 12:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____