
Midmark M11 - v1.0.5
Total Cycles: 371

Name: RITTER2
Sterilizer ID: V2313228
09/16/2021 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 165.3 F 2.9
4:00 186.3 F 7.3
6:00 208.2 F 12.7
8:00 227.2 F 8.1
10:00 237.2 F 12.5
12:00 241.8 F 11.6
14:00 239.9 F 10.5
16:00 253.8 F 17.4
18:00 265.4 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.1 F 28.8
2:00 272.1 F 29.3
3:00 271.6 F 28.4
4:00 272.4 F 29.3

Min 271.2 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.1 F 0.0
5:00 234.5 F 0.0
10:00 221.7 F 0.0
15:00 233.0 F 0.0
20:00 197.6 F 0.0
25:00 201.1 F 0.0
30:00 202.1 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:57:21
CYCLE COMPLETE

09/16/2021 10:57 AM
ACCEPTED: X
REJECTED:
Operator: U02