
Midmark M11 - v1.0.0
Total Cycles: 5227

Name: RITTER 1
Sterilizer ID: V2178198
01/08/2025 01:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.8 F 0.0
2:00 98.8 F 0.0
4:00 124.4 F 0.1
6:00 159.7 F 0.4
8:00 186.5 F 3.7
10:00 207.9 F 7.6
12:00 219.6 F 11.5
14:00 221.7 F 4.6
16:00 232.7 F 9.2
18:00 239.3 F 11.3
20:00 241.1 F 11.2
22:00 237.6 F 9.2
24:00 251.6 F 15.8
26:00 262.9 F 22.4
28:00 272.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 271.8 F 28.1
2:00 272.4 F 28.5
3:00 272.5 F 28.8
4:00 271.7 F 28.3

Min 271.5 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 206.0 F 0.0
10:00 193.5 F 0.0
15:00 192.8 F 0.0
20:00 189.0 F 0.0
25:00 188.6 F 0.0
30:00 188.7 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 28:10
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:06:25
CYCLE COMPLETE

01/08/2025 02:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____