

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3925

Name: RITTER2  
Sterilizer ID: V2313228  
09/10/2024 03:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.0 F 0.0  
2:00 173.9 F 2.5  
4:00 182.7 F 6.7  
6:00 205.3 F 11.4  
8:00 219.2 F 6.4  
10:00 231.5 F 11.7  
12:00 232.9 F 9.3  
14:00 237.8 F 9.8  
16:00 246.8 F 14.3  
18:00 258.9 F 20.9  
20:00 269.6 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 271.5 F 29.3  
2:00 271.6 F 29.3  
3:00 271.9 F 29.5  
4:00 272.0 F 29.8  
  
Min 271.2 F 29.3  
Max 272.4 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.6 F 0.0  
5:00 211.7 F 0.0  
10:00 221.7 F 0.0  
15:00 224.4 F 0.0  
20:00 223.1 F 0.0  
25:00 221.1 F 0.0  
30:00 219.2 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 20:55  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:00:14  
CYCLE COMPLETE

09/10/2024 04:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_