
Midmark M11 - v1.0.0
Total Cycles: 3336

Name: RITTER 1
Sterilizer ID: V2178198
06/05/2023 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.6 F 0.0
2:00 161.1 F 0.0
4:00 188.7 F 2.0
6:00 203.6 F 6.8
8:00 221.5 F 11.7
10:00 225.2 F 6.7
12:00 238.5 F 12.0
14:00 237.0 F 9.5
16:00 237.6 F 9.0
18:00 249.1 F 14.7
20:00 261.2 F 21.5
22:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.7
2:00 271.4 F 28.2
3:00 272.4 F 28.8
4:00 271.5 F 28.1

Min 271.4 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 212.3 F 0.0
10:00 224.3 F 0.0
15:00 230.8 F 0.0
20:00 220.9 F 0.0
25:00 218.9 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 22:29
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:40
CYCLE COMPLETE

06/05/2023 02:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____