

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1418

Name: RITTER 1  
Sterilizer ID: V2178198  
12/01/2021 03:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.3 F 0.0  
2:00 173.0 F 2.8  
4:00 198.4 F 7.1  
6:00 216.8 F 12.7  
8:00 226.0 F 6.8  
10:00 238.3 F 11.9  
12:00 236.3 F 9.0  
14:00 239.9 F 9.7  
16:00 248.0 F 14.0  
18:00 260.3 F 20.8  
20:00 271.3 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.4 F 28.7  
2:00 271.1 F 28.1  
3:00 272.3 F 28.9  
4:00 271.5 F 28.3

Min 271.0 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
5:00 238.5 F 0.0  
10:00 229.8 F 0.0  
15:00 229.0 F 0.0  
20:00 230.7 F 0.0  
25:00 228.9 F 0.0  
30:00 227.0 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 20:41  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:48  
CYCLE COMPLETE

12/01/2021 05:17 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02