
Midmark M11 - v1.0.0
Total Cycles: 1602

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2022 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.1 F 0.0
2:00 189.9 F 3.9
4:00 199.3 F 9.1
6:00 221.3 F 6.0
8:00 232.0 F 10.2
10:00 234.0 F 7.6
12:00 243.4 F 11.8
14:00 246.2 F 13.2
16:00 259.3 F 20.3
18:00 270.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 271.8 F 28.3
2:00 272.6 F 28.7
3:00 271.6 F 28.5
4:00 272.0 F 28.4

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 244.2 F 0.0
10:00 226.5 F 0.0
15:00 228.4 F 0.0
20:00 225.4 F 0.0
25:00 227.7 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 18:40
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

02/07/2022 03:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____