
Midmark M11 - v1.0.5
Total Cycles: 3028

Name: RITTER2
Sterilizer ID: V2313228
10/19/2023 08:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 80.8 F 0.0
4:00 98.0 F 0.4
6:00 109.0 F 2.4
8:00 137.7 F 5.2
10:00 165.0 F 8.3
12:00 218.4 F 12.2
14:00 209.4 F 4.8
16:00 225.3 F 8.6
18:00 233.8 F 12.8
20:00 233.3 F 8.7
22:00 241.2 F 11.8
24:00 243.2 F 12.2
26:00 255.1 F 18.4
28:00 265.8 F 25.0
30:00 271.9 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.3
1:00 272.8 F 29.8
2:00 272.1 F 29.8
3:00 271.7 F 29.2
4:00 272.8 F 29.7

Min 271.4 F 29.1
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 241.5 F 0.0
10:00 202.7 F 0.0
15:00 207.1 F 0.0
20:00 204.9 F 0.0
25:00 213.4 F 0.0
30:00 186.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:21
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:26
CYCLE COMPLETE

10/19/2023 09:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____