
Midmark M11 - v1.0.5
Total Cycles: 1964

Name: RITTER2
Sterilizer ID: V2313228
01/24/2023 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.8 F 0.0
2:00 172.9 F 0.8
4:00 183.9 F 5.2
6:00 205.6 F 10.4
8:00 228.3 F 6.0
10:00 237.0 F 11.9
12:00 238.2 F 10.2
14:00 237.0 F 9.2
16:00 251.6 F 16.4
18:00 263.7 F 23.5
20:00 271.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.2
1:00 272.4 F 29.2
2:00 272.4 F 29.3
3:00 272.5 F 29.4
4:00 272.6 F 29.6

Min 271.5 F 29.1
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 235.1 F 0.0
10:00 226.3 F 0.0
15:00 232.1 F 0.0
20:00 226.7 F 0.0
25:00 231.2 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

01/24/2023 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____