

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2137

Name: RITTER 1  
Sterilizer ID: V2178198  
07/15/2022 12:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.9 F 0.0  
2:00 150.4 F 0.3  
4:00 175.3 F 3.4  
6:00 191.6 F 7.5  
8:00 207.4 F 11.8  
10:00 220.0 F 5.0  
12:00 228.6 F 9.5  
14:00 238.7 F 10.4  
16:00 239.3 F 11.0  
18:00 236.5 F 8.6  
20:00 249.6 F 15.0  
22:00 261.0 F 21.3  
24:00 271.2 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.4 F 28.9  
2:00 271.4 F 28.3  
3:00 272.6 F 28.9  
4:00 271.4 F 28.2

Min 271.4 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 225.8 F 0.0  
10:00 236.5 F 0.0  
15:00 230.6 F 0.0  
20:00 225.9 F 0.0  
25:00 225.7 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 24:43  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:02:47  
CYCLE COMPLETE

07/15/2022 01:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_