
Midmark M11 - v1.0.5
Total Cycles: 2120

Name: RITTER2
Sterilizer ID: V2313228
03/06/2023 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.7 F 0.0
2:00 201.7 F 3.6
4:00 215.5 F 8.9
6:00 223.5 F 7.6
8:00 232.8 F 10.3
10:00 233.0 F 8.0
12:00 241.2 F 10.9
14:00 247.1 F 14.1
16:00 260.1 F 21.2
18:00 271.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.5
2:00 271.6 F 28.7
3:00 271.9 F 29.5
4:00 272.2 F 29.3

Min 271.2 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 228.5 F 0.0
10:00 232.5 F 0.0
15:00 228.9 F 0.0
20:00 224.2 F 0.0
25:00 225.8 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:44
CYCLE COMPLETE

03/06/2023 11:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____