

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2138

Name: RITTER2  
Sterilizer ID: V2313228  
03/13/2023 07:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.7 F 0.0  
2:00 77.2 F 0.0  
4:00 85.8 F 0.0  
6:00 107.5 F 0.0  
8:00 153.0 F 2.8  
10:00 172.7 F 6.4  
12:00 192.0 F 10.3  
14:00 217.4 F 6.9  
16:00 222.0 F 7.7  
18:00 233.4 F 11.9  
20:00 232.5 F 8.4  
22:00 242.8 F 12.4  
24:00 242.5 F 12.2  
26:00 254.0 F 18.1  
28:00 264.5 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 271.8 F 29.2  
2:00 272.1 F 29.4  
3:00 272.1 F 29.7  
4:00 272.2 F 29.9

Min 271.5 F 29.1  
Max 272.4 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 232.7 F 0.0  
10:00 213.0 F 0.0  
15:00 214.9 F 0.0  
20:00 208.1 F 0.0  
25:00 214.1 F 0.0  
30:00 216.7 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 29:56  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:07:57  
CYCLE COMPLETE

03/13/2023 08:19 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_