

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1143

Name: RITTER 1  
Sterilizer ID: V2178198  
09/02/2021 03:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 219.1 F 0.0  
2:00 199.0 F 3.1  
4:00 222.8 F 8.1  
6:00 220.6 F 13.5  
8:00 233.7 F 8.9  
10:00 241.1 F 10.6  
12:00 244.0 F 12.2  
14:00 241.4 F 10.8  
16:00 254.6 F 17.5  
18:00 265.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.7 F 29.0  
2:00 271.6 F 28.3  
3:00 271.8 F 28.9  
4:00 272.5 F 28.9  
  
Min 271.0 F 28.2  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.1 F 0.0  
5:00 231.8 F 0.0  
10:00 230.1 F 0.0  
15:00 218.0 F 0.0  
20:00 222.9 F 0.0  
25:00 225.7 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 3:31  
HEATING: 19:44  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:27  
CYCLE COMPLETE

09/02/2021 04:26 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02