
Midmark M11 - v1.0.0
Total Cycles: 4463

Name: RITTER 1
Sterilizer ID: V2178198
04/16/2024 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.4 F 0.0
2:00 179.5 F 3.2
4:00 211.4 F 7.2
6:00 216.0 F 12.4
8:00 224.2 F 6.3
10:00 236.4 F 11.1
12:00 232.7 F 7.7
14:00 245.7 F 12.6
16:00 244.5 F 12.3
18:00 257.1 F 18.9
20:00 268.2 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 272.6 F 28.8
2:00 271.8 F 28.0
3:00 272.5 F 28.8
4:00 272.1 F 28.3

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 225.3 F 0.0
10:00 216.8 F 0.0
15:00 221.7 F 0.0
20:00 214.3 F 0.0
25:00 218.7 F 0.0
30:00 211.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

04/16/2024 12:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____