
Midmark M11 - v1.0.0
Total Cycles: 744

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2021 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.6 F 0.0
2:00 187.6 F 3.3
4:00 205.5 F 8.1
6:00 232.7 F 8.2
8:00 230.6 F 10.5
10:00 232.3 F 7.7
12:00 240.0 F 10.6
14:00 247.0 F 13.4
16:00 260.0 F 20.5
18:00 271.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.9 F 28.8
2:00 271.8 F 28.9
3:00 271.7 F 28.8
4:00 271.2 F 28.7

Min 271.1 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 239.3 F 0.0
10:00 228.9 F 0.0
15:00 228.7 F 0.0
20:00 190.2 F 0.0
25:00 207.7 F 0.0
30:00 182.7 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 19:15
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

03/24/2021 03:24 PM
ACCEPTED: X
REJECTED:
Operator: U01