
Midmark M11 - v1.0.5
Total Cycles: 2067

Name: RITTER2
Sterilizer ID: V2313228
02/17/2023 01:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.7 F 0.0
2:00 150.0 F 0.0
4:00 160.0 F 0.9
6:00 186.1 F 4.9
8:00 208.9 F 9.1
10:00 220.9 F 9.1
12:00 233.4 F 10.3
14:00 232.9 F 8.2
16:00 241.6 F 11.6
18:00 245.7 F 13.5
20:00 258.2 F 20.3
22:00 269.3 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.4 F 29.4
2:00 272.0 F 29.6
3:00 272.1 F 29.2
4:00 272.1 F 29.7

Min 271.6 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 223.1 F 0.0
10:00 220.7 F 0.0
15:00 219.1 F 0.0
20:00 223.8 F 0.0
25:00 224.6 F 0.0
30:00 233.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:06
CYCLE COMPLETE

02/17/2023 02:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____