
Midmark M11 - v1.0.0
Total Cycles: 4947

Name: RITTER 1
Sterilizer ID: V2178198
10/01/2024 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.2 F 0.0
2:00 152.5 F 0.2
4:00 172.2 F 1.5
6:00 192.8 F 5.1
8:00 205.9 F 8.6
10:00 221.4 F 12.7
12:00 226.2 F 7.5
14:00 236.2 F 12.2
16:00 235.6 F 9.3
18:00 239.4 F 9.9
20:00 246.3 F 13.4
22:00 257.5 F 19.5
24:00 267.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 271.9 F 29.1
2:00 272.6 F 29.3
3:00 272.3 F 29.0
4:00 271.7 F 28.6

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 215.0 F 0.0
10:00 212.9 F 0.0
15:00 223.3 F 0.0
20:00 230.1 F 0.0
25:00 232.3 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 25:35
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:04:13
CYCLE COMPLETE

10/01/2024 12:40 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____